



Product Service

EC-Attestation of Compliance

No. E8N 08 06 62545 004

Holder of Certificate: DA TECH CO., Ltd
#710-C, Woolim Lionsvalley, 425
Cheongcheon-dong, Bupyeong-gu, Incheon 403-858
REPUBLIC OF KOREA

Name of Object: Electrical equ. for measurement, control and
laboratory use
Breathalyzer

Model(s): DA-8000

Description of Object: Rated input voltage: 3 V d.c.
(1.5 V Battery * 2EA)

Tested according to: EN 61326:1997+A1:1998+A2:2001+A3:2003 Class B

This EC-Attestation of Compliance is issued according to the Directive 2004/108/EC relating to electromagnetic compatibility on a voluntary basis. It confirms that the listed apparatus complies with all essential requirements of the EMC directive and applies only to the sample and its technical documentation submitted to TÜV SÜD Product Service GmbH for testing and certification. See also notes overleaf.

Test report no.: ESTCE0805-005

Date, 2008-06-18



After preparation of the necessary technical documentation as well as the conformity declaration the required CE marking can be affixed on the product. Other relevant directives have to be observed.

Page 1 of 1